

Donnelly, Margaret

From: Derboghossian, Michael
Sent: June 01, 2007 4:02 PM
To: Andry, Chris
Cc: Donnelly, Margaret
Subject: Biosquare 3 Equipment

Chris,

Equipment in the laboratories on the third floor of Biosquare 3 was tested for electrical safety on March 4, 2007. In addition to electrical safety, centrifuges were rate-tested using a tachometer to verify RPMs. There were no failures to report. The next scheduled testing will occur in March 2008.

After I performed testing, I forwarded you a report from our database indicating the inventory of devices on the third floor. If you need additional copies, please let me know.

Thanks,
Mike

Michael DerBoghossian

Clinical Engineer
Boston Medical Center
T 617-414-4692
F 617-414-4694
P 617-638-5795, ID# 6039
E michael.derboghossian@bmc.org